

AOFSRR 2015

Asia Oceania Forum for Synchrotron
Radiation Research



in conjunction with

**USER
MEETING
2015**

National Centre for Synchrotron Science

25-27 NOVEMBER 2015

Contribution ID : 168

Type : Poster

A Comparison of the Measurement of Protein Solutions on a MetalJet Equipped Laboratory SAXS Instrument and a Synchrotron Beamline

Friday, 27 November 2015 13:30 (45)

SFPQ is a protein that suppresses tumours in humans. Two samples were prepared from the same batch and frozen until just prior to measurement. Traditionally the investigation of protein solutions has been limited to synchrotron facilities, though more recently instruments making use of microfocus rotating anodes have allowed more measurements to be made in the laboratory. One sample was measured on a Bruker AXS Nanostar using the MetalJet X-Ray source from Excillum, the latest generation of x-ray generators for laboratories. The MetalJet X-Ray source makes use of a pressurised liquid metal anode capable of withstanding a high power electron beam and the resulting x-ray beam is much brighter than conventional sources. The SAXS instrument was further equipped with long Montel multi-layer mirrors and home-built "scatterless" slits. While rotating anode sources often require at least 30 minutes of collection time for scattering curves from weakly scattering biomolecular solutions the MetalJet data presented here were collected in 1 minute. While this is still significantly longer than what is required at synchrotron facilities it represents a major improvement on current generation laboratory instruments. The new SAXS facilities in Western Australia will provide the user community with access to a larger range of experiments and also help in generating preliminary data for experiments that specifically require synchrotron access.

Keywords

Primary author(s) : Dr JARRETT, Kevin (Curtin University)

Co-author(s) : Prof. BOND, Charlie (University of Western Australia); Prof. BUCKLEY, Craig (Curtin University); Mr KNOTT, Gavin (University of Western Australia); LYNGSO, J (Interdisciplinary Nanoscience Center (iNANO) and Department of Chemistry, University of Aarhus, DK-8000 Aarhus, Denmark); PEDERSON, J. S. (Interdisciplinary Nanoscience Center (iNANO) and Department of Chemistry, University of Aarhus, DK-8000 Aarhus, Denmark)

Presenter(s) : Dr ROWLES, Matthew (Curtin University)

Session Classification : Poster Session 2

Track Classification : Structural Biology