

AOFSRR 2015

Asia Oceania Forum for Synchrotron
Radiation Research



ANSTO

Australian
Synchrotron



in conjunction with

**USER
MEETING
2015**

National Centre for Synchrotron Science

25-27 NOVEMBER 2015

Contribution ID : 200

Type : Poster

Chemical Speciation Imaging at Environmentally Relevant Concentrations using X-ray Fluorescence Microscopy

Friday, 27 November 2015 13:30 (45)

X-ray fluorescence microscopy (XFM) can be used for elemental and chemical microanalysis across length scales ranging from millimeter to nanometer. XFM is ideally suited to quantitatively map trace elements within whole plant and other biological specimens, environmental and soil samples. The elemental sensitivity of the X-ray fluorescence probe provides valuable information in a diversity of environmental sciences, and the high penetration of hard X-rays enables measurement of whole cells, tissue sections and a diverse range of environmental samples with a minimum of preparation.

Rapid advances in X-ray fluorescence detection methods such as the Maia detector now enable high definition images approaching megapixel per minute rates. The ability to rapidly acquire 2D images enables 3D information such as fluorescence tomography to be obtained in realistic times. Chemical speciation (or valence) imaging (CSI) is a technique where the third dimension is spectroscopic detail. CSI results in an X-ray Absorption Near Edge Structure (XANES) spectra from the X-ray fluorescence signal at each pixel in the spatial image.

CSI has been demonstrated at the Australian Synchrotron XFM beamline with micron resolution and moderate definition (10K pixels) across a diverse range of sciences and applications from environmental chemistry to arsenic toxicity in crop production. Studies probing and optimising the efficiency and sensitivity of CSI to achieve measurements at environmentally relevant concentrations will be presented.

Keywords

X-ray fluorescence microscopy, chemical speciation imaging, environmentally relevant concentrations

Primary author(s) : PATERSON, David (Australian Synchrotron)

Co-author(s) : ETSCHMANN, Barbara (Monash University); Dr RYAN, Chris (CSIRO); HOWARD, Daryl (Australian Synchrotron); Dr LOMBI, Enzo (University of South Australia); Dr SPIERS, Kathryn (Australian Synchrotron); Dr DE JONGE, Martin (Australian Synchrotron); Dr KOPITTKE, Peter (The University of Queensland)

Presenter(s) : PATERSON, David (Australian Synchrotron)

Session Classification : Poster Session 2

Track Classification : Beamlines, Instrumentation and Techniques